



Biochrom Ltd
Certificate No. 890333

CERTIFICATE OF ANALYSIS

for

EZ Nin Reagent

Part Number: 80-6000-12

Control Number: 20790

Quantity: 2.00 litres

The above product gave the following results on a standard Lithium High-Performance System:

Baseline drift: 0.03 $\Delta mV/\Delta t$ at 440nm

Baseline drift: <0.01 $\Delta mV/\Delta t$ at 570nm

Drift Specification <0.09 at 440nm, <0.06 at 570nm

Expiry Date: Mar 2015

Chemical QC

Date: 31 Mar 2014

| | | | |
|--|---|--------------------|------------------------|
| Postal Address | Telephone | Telefax | Email: |
| Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England | +44 (0)1223 423723 | +44 (0)1223 420164 | support@biochrom.co.uk |
| | Web Site: http://www.biochrom.co.uk | | |

Registered in England No: 3526954 Registered Office: 22 Cambridge Science Park, Milton Road, Cambridge, CB4 0FJ, England.