

CERTIFICATE OF ANALYSIS

for

Sodium Hydrolysate Buffer 5

Part Number: 80-2038-09

Control Number: 20878

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 6.470 @ 20.0°C

pH Specification = 6.45 ± 0.02 @ 20.0°C

Conductivity = 77.2 mScm⁻¹ @ 20.0°C

Conductivity Specification = 78.8 ± 3.6 mScm⁻¹ @ 20.0°C

Chemical QC

Date: 15 October 2014

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		