



Biochrom Ltd
Certificate No. 890333

CERTIFICATE OF ANALYSIS

for

Sodium Oxidized Buffer 1

Part Number: 80-2037-80

Control Number: 20914

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 3.349 @ 20.0°C

pH Specification = 3.35 ± 0.02 @ 20.0°C

Conductivity = 15.4 mScm⁻¹ @ 20.0°C

Conductivity Specification = 16.4 ± 2.8 mScm⁻¹ @ 20.0°C

Chemical QC

Date: 21 August 2015

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 22 Cambridge Science Park, Milton Road, Cambridge, CB4 0FJ, England.