



Biochrom Ltd  
Certificate No. 890333

## CERTIFICATE OF ANALYSIS

for

### EZ Nin Reagent

Part Number: 80-6000-12

Control Number: 20953

Quantity: 2.00 litres

The above product gave the following results on a standard Lithium High-Performance System:

Baseline drift: 0.04  $\Delta mV/\Delta t$  at 440nm

Baseline drift: 0.02  $\Delta mV/\Delta t$  at 570nm

Drift Specification <0.09 at 440nm, <0.06 at 570nm

Expiry Date: Jun 2015

Chemical QC

Date: 6 Oct 2014

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: 22 Cambridge Science Park, Milton Road, Cambridge, CB4 0FJ, England.