



Biochrom Ltd
Certificate No. 890333

CERTIFICATE OF ANALYSIS

for

Sodium Regeneration Buffer 6

Part Number: 80-2037-57

Control Number: 21130

Quantity: 1.00 litre

The above product gave the following results on analysis:

Basic

Conductivity = 70.4 mScm⁻¹ @ 20.0°C

Conductivity Specification = 72.1 ± 3.6 mScm⁻¹ @ 20.0°C

Chemical QC

Date: 28 November 2014

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 22 Cambridge Science Park, Milton Road, Cambridge, CB4 0FJ, England.