

## CERTIFICATE OF ANALYSIS

for

### Lithium Regeneration Buffer 6

Part Number: 80-2038-20

Control Number: 21174

Quantity: 1.00 litre

The above product gave the following results on analysis:

Basic

Conductivity = 54.8 mScm<sup>-1</sup> @ 20.0°C

Conductivity Specification = 52.3 ± 2.7 mScm<sup>-1</sup> @ 20.0°C

Chemical QC

Date: 19 January 2015

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		