

CERTIFICATE OF ANALYSIS

for

Lithium Buffer 1

Part Number: 80-2038-15

Control Number: 21201

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 2.802 @ 20.0°C

pH Specification = 2.80 ± 0.02 @ 20.0°C

Conductivity = 15.2 mScm⁻¹ @ 20.0°C

Conductivity Specification = 15.2 ± 0.8 mScm⁻¹ @ 20.0°C

Chemical QC

Date: 13 December 2014

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		