



Biochrom Ltd
Certificate No. 890333

CERTIFICATE OF ANALYSIS

for

Ultrosolve Plus

Part Number: 80-2117-65

Control Number: 21227

Quantity: 1.75 litres

The above product gave the following results on analysis:

pH = 5.670 @ 20.0°C

pH Specification = 5.65 ± 0.1 @ 20.0°C

Conductivity = 20.8 mScm⁻¹ @ 20.0°C

Conductivity Specification = 20.5 ± 3 mScm⁻¹ @ 20.0°C

Chemical QC

Date: 18 February 2015

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 22 Cambridge Science Park, Milton Road, Cambridge, CB4 0FJ, England.