

CERTIFICATE OF ANALYSIS

for

Sodium Oxidized Buffer 1

Part Number: 80-2037-80

Control Number: 21251

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 3.350 @ 20.0°C

pH Specification = 3.35 ± 0.02 @ 20.0°C

Conductivity = 16.2 mScm⁻¹ @ 20.0°C

Conductivity Specification = 16.4 ± 2.8 mScm⁻¹ @ 20.0°C

Chemical QC

Date: 24 April 2015

Postal Address

Biochrom Ltd
22 Cambridge Science Park
Milton Road
Cambridge CB4 0FJ
England

Telephone

+44 (0)1223 423723

Telefax

+44 (0)1223 420164

Email:

support@biochrom.co.uk

Web Site: <http://www.biochrom.co.uk>