

## CERTIFICATE OF ANALYSIS

for

### Sodium Accelerated Buffer 2

Part Number: 80-2115-22

Control Number: 21276

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 3.507 @ 20.0°C

pH Specification = 3.50 ± 0.02 @ 20.0°C

Conductivity = 11.2 mScm-1 @ 20.0°C

Conductivity Specification = 10.5 ± 1.0 mScm-1 @ 20.0°C

Chemical QC

Date: 17 December 2015

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		