

CERTIFICATE OF ANALYSIS

for

Lithium Buffer 2

Part Number: 80-2038-16

Control Number: 21281

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 2.996 @ 20.0°C

pH Specification = 3.00 ± 0.02 @ 20.0°C

Conductivity = 21.6 mScm-1 @ 20.0°C

Conductivity Specification = 21.0 ± 1.2 mScm-1 @ 20.0°C

Chemical QC

Date: 06 February 2015

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		