

## CERTIFICATE OF ANALYSIS

for

### Ultrosolve Plus

Part Number: 80-2117-65

Control Number: 21282

Quantity: 1.75 litres

The above product gave the following results on analysis:

pH = 5.655 @ 20.0°C

pH Specification = 5.65 ± 0.1 @ 20.0°C

Conductivity = 18.8 mScm<sup>-1</sup> @ 20.0°C

Conductivity Specification = 20.5 ± 3 mScm<sup>-1</sup> @ 20.0°C

Chemical QC

Date: 04 March 2015

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		