

## CERTIFICATE OF ANALYSIS

for

### Sodium Accelerated Buffer 3

Part Number: 80-2115-23

Control Number: 21290

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 4.407 @ 20.0°C

pH Specification = 4.40 ± 0.02 @ 20.0°C

Conductivity = 9.59 mScm-1 @ 20.0°C

Conductivity Specification = 9.0 ± 1.0 mScm-1 @ 20.0°C

Chemical QC

Date: 17 December 2015

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		