

CERTIFICATE OF ANALYSIS

for

Sodium Oxidized Buffer 4

Part Number: 80-2037-88

Control Number: 21304

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 8.593 @ 20.0°C

pH Specification = 8.60 ± 0.02 @ 20.0°C

Conductivity = 36.0 mScm-1 @ 20.0°C

Conductivity Specification = 38.6 ± 2.8 mScm-1 @ 20.0°C

Chemical QC

Date: 17 April 2015

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		