

CERTIFICATE OF ANALYSIS

for

Ultra Ninhydrin Solution

Part Number: 80-2117-64

Control Number: 21345

Quantity: 250mL

The above product gave the following results on analysis:

Base line drift no more than:

0.07 $\Delta mV/\Delta t$ at 440nm

0.04 $\Delta mV/\Delta t$ at 570nm

Colorimetric test:

Absorption at 485nm [Neat, ref. to water. pass if < 0.10] = -0.033

Absorption at 600nm [Neat, ref. to water. pass if < 0.05] = -0.038

Chemical QC

Date: 21 April 2015

| | | | |
|--|---|--------------------|------------------------|
| Postal Address | Telephone | Telefax | Email: |
| Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England | +44 (0)1223 423723 | +44 (0)1223 420164 | support@biochrom.co.uk |
| | Web Site: http://www.biochrom.co.uk | | |