



Biochrom Ltd
Certificate No. 890333

CERTIFICATE OF ANALYSIS

for

Sodium Accelerated Buffer 4

Part Number: 80-2115-24

Control Number: 21359

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 9.218 @ 20.0°C

pH Specification = 9.23 ± 0.02 @ 20.0°C

Conductivity = 48.7 mScm⁻¹ @ 20.0°C

Conductivity Specification = 49.0 ± 2.0 mScm⁻¹ @ 20.0°C

Chemical QC

Date: 18 December 2015

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.