

## CERTIFICATE OF ANALYSIS

for

### Ultrosolve Plus

Part Number: 80-2117-65

Control Number: 21385

Quantity: 1.75 litres

The above product gave the following results on analysis:

pH = 5.641 @ 20.0°C

pH Specification = 5.65 ± 0.1 @ 20.0°C

Conductivity = 19.6 mScm<sup>-1</sup> @ 20.0°C

Conductivity Specification = 20.5 ± 3 mScm<sup>-1</sup> @ 20.0°C

Chemical QC

Date: 19 May 2015

---

|  |   |                    |                        |
|--|---|--------------------|------------------------|
| Postal Address   | Telephone   | Telefax            | Email:                 |
| Biochrom Ltd<br>22 Cambridge Science Park<br>Milton Road<br>Cambridge CB4 0FJ<br>England | +44 (0)1223 423723  | +44 (0)1223 420164 | support@biochrom.co.uk |
|  | Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a> |                    |                        |