

## CERTIFICATE OF ANALYSIS

for

### Sodium Hydro/Oxid Buffer 2

Part Number: 80-2038-08

Control Number: 21405

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 4.239 @ 20.0°C

pH Specification = 4.25 ± 0.02 @ 20.0°C

Conductivity = 14.3 mScm-1 @ 20.0°C

Conductivity Specification = 14.2 ± 1.1 mScm-1 @ 20.0°C

Chemical QC

Date: 18 August 2015

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		