

## CERTIFICATE OF ANALYSIS

for

### Ultra Ninhydrin Solution

Part Number: 80-2117-64

Control Number: 21455

Quantity: 250mL

The above product gave the following results on analysis:

Base line drift no more than:

0.08  $\Delta mV/\Delta t$  at 440nm

<0.01  $\Delta mV/\Delta t$  at 570nm

Colorimetric test:

Absorption at 485nm [Neat, ref. to water. pass if <0.10] = -0.032

Absorption at 600nm [Neat, ref. to water. pass if <0.05] = 0.056

Chemical QC

Date: 02 June 2015

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		