



Biochrom Ltd
Certificate No. 890333

CERTIFICATE OF ANALYSIS

for

Sodium Regeneration Buffer 6

Part Number: 80-2037-57

Control Number: 21470

Quantity: 1.00 litre

The above product gave the following results on analysis:

Basic

Conductivity = 70.4 mScm⁻¹ @ 20.0°C

Conductivity Specification = 72.1 ± 3.6 mScm⁻¹ @ 20.0°C

Chemical QC

Date: 23 November 2015

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.