

## CERTIFICATE OF ANALYSIS

for

### Sodium Hydrolysate Buffer 1

Part Number: 80-2037-68

Control Number: 21509

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 3.195 @ 20.0°C

pH Specification = 3.20 ± 0.02 @ 20.0°C

Conductivity = 15.2 mScm<sup>-1</sup> @ 20.0°C

Conductivity Specification = 15.2 ± 0.8 mScm<sup>-1</sup> @ 20.0°C

Chemical QC

Date: 06 November 2015

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		