

CERTIFICATE OF ANALYSIS

for

Ultra Ninhydrin Solution

Part Number: 80-2117-64

Control Number: 21515

Quantity: 250mL

The above product gave the following results on analysis:

Base line drift no more than:

0.07 $\Delta mV/\Delta t$ at 440nm

<0.01 $\Delta mV/\Delta t$ at 570nm

Colorimetric test:

Absorption at 485nm [Neat, ref. to water. pass if<0.10] = -0.014

Absorption at 600nm [Neat, ref. to water. pass if<0.05] = -0.048

Chemical QC

Date: 21 July 2015

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		