

## CERTIFICATE OF ANALYSIS

for

### Lithium Buffer 5

Part Number: 80-2037-69

Control Number: 21631

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 3.535 @ 20.0°C

pH Specification = 3.55 ± 0.02 @ 20.0°C

Conductivity = 84.3 mScm<sup>-1</sup> @ 20.0°C

Conductivity Specification = 86.2 ± 3.4 mScm<sup>-1</sup> @ 20.0°C

Chemical QC

Date: 27 October 2015

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 22 Cambridge Science Park Milton Road Cambridge CB4 0FJ England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		