



Biochrom Ltd
Certificate No. 890333

CERTIFICATE OF ANALYSIS

for

Sodium Oxidized Buffer 1

Part Number: 80-2037-80

Control Number: 21636

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 3.354 @ 20.0°C

pH Specification = 3.35 ± 0.02 @ 20.0°C

Conductivity = 17.70 mScm⁻¹ @ 20.0°C

Conductivity Specification = 16.4 ± 2.8 mScm⁻¹ @ 20.0°C

Chemical QC

Date: 23 October 2015

Postal Address

Biochrom Ltd
1020 Cambourne Business Park
Cambourne
Cambridge CB23 6DW
England

Telephone

+44 (0)1223 423723

Telefax

+44 (0)1223 420164

Email:

support@biochrom.co.uk

Web Site: <http://www.biochrom.co.uk>

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.