



Biochrom Ltd
Certificate No. 890333

CERTIFICATE OF ANALYSIS

for

Ultrosolve Plus

Part Number: 80-2117-65

Control Number: 21638

Quantity: 1.75 litres

The above product gave the following results on analysis:

pH = 5.632 @ 20.0 °C

pH Specification = 5.65 ± 0.1 @ 20 °C

Conductivity = 19.91 mScm⁻¹ @ 20 °C

Conductivity Specification = 20.5 ± 3.0 mScm⁻¹ @ 20.0 °C

Chemical QC

Date: 29 March 2016

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.