

CERTIFICATE OF ANALYSIS

for

Ultra Ninhydrin Solution

Part Number: 80-2117-64

Control Number: 21641

Quantity: 250mL

The above product gave the following results on analysis:

Base line drift no more than:

0.03 $\Delta mV/\Delta t$ at 440nm

0.02 $\Delta mV/\Delta t$ at 570nm

Colorimetric test:

Absorption at 485nm [Neat, ref. to water. pass if < 0.10] = 0.041

Absorption at 600nm [Neat, ref. to water. pass if < 0.05] = 0.010

Chemical QC

Date: 08 January 2016

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		