



Biochrom Ltd  
Certificate No. 890333

## CERTIFICATE OF ANALYSIS

for

### Polyamine Buffer

Part Number: 80-2037-74

Control Number: 21649

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 5.605 @ 20.0°C

pH Specification = 5.60 ± 0.02 @ 20.0°C

Conductivity = 87.5 mScm-1 @ 20.0°C

Conductivity Specification = 89.0 ± 5.4 mScm-1 @ 20.0°C

Chemical QC

Date: 19 February 2016

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.