



Biochrom Ltd
Certificate No. 890333

CERTIFICATE OF ANALYSIS

for

Sodium Accelerated Buffer 2

Part Number: 80-2115-22

Control Number: 21651

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 3.503 @ 20.0 °C

pH Specification = 3.50 ± 0.02 @ 20 °C

Conductivity = 10.70 mScm⁻¹ @ 20 °C

Conductivity Specification = 10.5 ± 1.0 mScm⁻¹ @ 20.0 °C

Chemical QC

Date: 29 Mar 2016

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.