



Biochrom Ltd  
Certificate No. 890333

## CERTIFICATE OF ANALYSIS

for

### Sodium Regeneration Buffer 6

Part Number: 80-2037-57

Control Number: 21655

Quantity: 1.00 litre

The above product gave the following results on analysis:

Basic

Conductivity = 73.4 mScm-1 @ 20 °C

Conductivity Specification = 72.1 ± 3.6 mScm-1 @ 20.0 °C

Chemical QC

Date: 07 March 2016

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.