



Biochrom Ltd  
Certificate No. 890333

## CERTIFICATE OF ANALYSIS

for

### Ultrosolve Plus

Part Number: 80-2117-65

Control Number: 21658

Quantity: 1.75 litres

The above product gave the following results on analysis:

pH = 5.643 @ 20.0°C

pH Specification = 5.65 ± 0.1 @ 20.0°C

Conductivity = 20.25 mScm<sup>-1</sup> @ 20.0°C

Conductivity Specification = 20.5 ± 3 mScm<sup>-1</sup> @ 20.0°C

Chemical QC

Date: 03 April 2016

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.