

CERTIFICATE OF ANALYSIS

for

Lithium Buffer 2

Part Number: 80-2038-16

Control Number: 21659

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 3.016 @ 20.0°C

pH Specification = 3.00 ± 0.02 @ 20.0°C

Conductivity = 20.9 mScm-1 @ 20.0°C

Conductivity Specification = 21.0 ± 1.2 mScm-1 @ 20.0°C

Chemical QC

Date: 18 March 2016

| | | | |
|--|---|--------------------|------------------------|
| Postal Address | Telephone | Telefax | Email: |
| Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England | +44 (0)1223 423723 | +44 (0)1223 420164 | support@biochrom.co.uk |
| | Web Site: http://www.biochrom.co.uk | | |