

CERTIFICATE OF ANALYSIS

for

Lithium Loading Buffer

Part Number: 80-2038-10

Control Number: 21670

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 2.192 @ 20.0°C

pH Specification = 2.20 ± 0.02 @ 20.0°C

Conductivity = 17.3 mScm-1 @ 20.0°C

Conductivity Specification = 17.2 ± 1.1 mScm-1 @ 20.0°C

Chemical QC

Date: 24 May 2016

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		