



Biochrom Ltd  
Certificate No. 890333

## CERTIFICATE OF ANALYSIS

for

### EZNin Reagent

Part Number: 80-6000-12

Control Number: 21671

Quantity: 2.0 Litres

The above product gave the following results on analysis:

Base line drift:

$\Delta mV/\Delta t$  on 440nm channel [Pass if  $<0.09$  mV/min] = 0.045 mV/min

$\Delta mV/\Delta t$  on 570nm channel [Pass if  $<0.06$  mV/min] = 0.005 mV/min

Chemical QC

Date: 11<sup>th</sup> May 2016

|  |   |                    |                        |
|--|---|--------------------|------------------------|
| Postal Address   | Telephone   | Telefax            | Email:                 |
| Biochrom Ltd<br>1020 Cambourne Business Park<br>Cambourne<br>Cambridge CB23 6DW<br>England | +44 (0)1223 423723  | +44 (0)1223 420164 | support@biochrom.co.uk |
|  | Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a> |                    |                        |

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.