

CERTIFICATE OF ANALYSIS

For

Sodium Accelerated Buffer 2

Part Number: 80-2115-22

Batch (Control) Number: 21673

Quantity: 2 Litres

Expiry Date: May 2019

The above product gave the following results on analysis:

pH = 3.488 @ 20.0°C

pH Specification = 3.48 - 3.52 @ 20.0°C

Conductivity = 11.1 mScm⁻¹ @ 20.0°C

Conductivity Specification = 9.5 - 11.5 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
Daniel Garrod

Date of Issue: 23 May 2016

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.