

CERTIFICATE OF ANALYSIS

for

Lithium Buffer 3

Part Number: 80-2099-83

Control Number: 21679

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 3.140 @ 20.0°C

pH Specification = 3.15 ± 0.02 @ 20.0°C

Conductivity = 33.9 mScm-1 @ 20.0°C

Conductivity Specification = 36.6 ± 3.2 mScm-1 @ 20.0°C

Chemical QC

Date: 22 July 2016

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		