



Biochrom Ltd  
Certificate No. 890333

## CERTIFICATE OF ANALYSIS

for

### Sodium Hydrolysate Buffer 1

Part Number: 80-2037-68

Control Number: 21682

Quantity: 2.0 Litres

The above product gave the following results on analysis:

pH = 3.204 @ 20.0 °C

pH Specification = 3.20 ± 0.02 @ 20.0 °C

Conductivity = 16.0 mScm<sup>-1</sup> @ 20.0 °C

Conductivity Specification = 15.2 ± 0.8 mScm<sup>-1</sup> @ 20.0 °C

Chemical QC

Date: 08 July 2016

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.