



Biochrom Ltd
Certificate No. 890333

CERTIFICATE OF ANALYSIS

for

Lithium Regeneration Buffer 6

Part Number: 80-2038-20

Control Number: 21683

Quantity: 1.00 litre

The above product gave the following results on analysis:

Basic

Conductivity = 50.1 mScm-1 @ 20.0°C

Conductivity Specification = 52.3 ± 2.7 mScm-1 @ 20.0°C

Chemical QC

Date: 04 August 2016

| | | | |
|--|---|--------------------|------------------------|
| Postal Address | Telephone | Telefax | Email: |
| Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England | +44 (0)1223 423723 | +44 (0)1223 420164 | support@biochrom.co.uk |
| | Web Site: http://www.biochrom.co.uk | | |

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.