

CERTIFICATE OF ANALYSIS

for

Lithium Buffer 2

Part Number: 80-2038-16

Control Number: 21693

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 2.999 @ 20.0°C

pH Specification = 3.00 ± 0.02 @ 20.0°C

Conductivity = 21.9 mScm⁻¹ @ 20.0°C

Conductivity Specification = 21.0 ± 1.2 mScm⁻¹ @ 20.0°C

Chemical QC

Date: 01 February 2017

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		