

CERTIFICATE OF ANALYSIS

for

Lithium Buffer 5

Part Number: 80-2037-69

Control Number: 21695

Quantity: 2.00 litres

The above product gave the following results on analysis:

pH = 3.530 @ 20.0°C

pH Specification = 3.55 ± 0.02 @ 20.0°C

Conductivity = 87.3 mScm-1 @ 20.0°C

Conductivity Specification = 86.2 ± 3.4 mScm-1 @ 20.0°C

Chemical QC

Date: 19 December 2016

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		