



# CERTIFICATE OF ANALYSIS

For

## Lithium Buffer 3

Part Number: 80-2099-83

Batch (Control) Number: 21696

Quantity: 2 Litres

The above product gave the following results on analysis:

pH = 3.15 @20.0°C

pH Specification = 3.13 – 3.17 @20.0°C

Conductivity = 33.70 mScm<sup>-1</sup> @20.0°C

Conductivity Specification = 33.41 – 39.71mScm<sup>-1</sup>

Chemical QC Ruth Palmer Date: 16/12/2016

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.