

CERTIFICATE OF ANALYSIS

For

Lithium Buffer 1

Part Number: 80-2038-15

Batch (Control) Number: 900001-4

Quantity: 2 Litres

Expiry Date: October 2020

The above product gave the following results on analysis:

pH = 2.80 @ 20.0°C

pH Specification = 2.78 - 2.82 @ 20.0°C

Conductivity = 14.7 mScm⁻¹ @ 20.0°C

Conductivity Specification = 14.4 - 16.0 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
Ruth Palmer

Date of Issue: 13 October 2017

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.