

CERTIFICATE OF ANALYSIS

For

Lithium Buffer 2

Part Number: 80-2038-16

Batch (Control) Number: 900002-3

Quantity: 2 Litres

Expiry Date: March 2021

The above product gave the following results on analysis:

pH = 3.00 @ 20.0°C

pH Specification = 2.98 - 3.02 @ 20.0°C

Conductivity = 21.0 mScm⁻¹ @ 20.0°C

Conductivity Specification = 19.8 - 22.2 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
Ruth Palmer

Date of Issue: 27 March 2018

| | | | |
|--|---|--------------------|------------------------|
| Postal Address | Telephone | Telefax | Email: |
| Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England | +44 (0)1223 423723 | +44 (0)1223 420164 | support@biochrom.co.uk |
| | Web Site: http://www.biochrom.co.uk | | |

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.