

CERTIFICATE OF ANALYSIS

For

Lithium Loading Buffer NC

Part Number: 80-6003-07

Batch (Control) Number: 900006-6NC

Quantity: 2 Litres

Expiry Date: March 2022

The above product gave the following results on analysis:

pH = 2.21 @ 20.0°C

pH Specification = 2.18 - 2.22 @ 20.0°C

Conductivity = 17.0 mScm⁻¹ @ 20.0°C

Conductivity Specification = 16.1 - 18.3 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
Daniel Garrod

Date of Issue: 29 March 2019

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.