

CERTIFICATE OF ANALYSIS

For

Lithium Regeneration Buffer 6

Part Number: 80-2038-20

Batch (Control) Number: 900007-4

Quantity: 1 Litre

Expiry Date: May 2021

The above product gave the following results on analysis:

Conductivity = 54.3 mScm⁻¹ @ 20.0°C

Conductivity Specification = 49.6 - 55.0 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
Ruth Palmer

Date of Issue: 11 May 2018

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.