



Biochrom Ltd
Certificate No. 890333

CERTIFICATE OF ANALYSIS

for

EZ Nin Reagent

Part Number: 80-6000-12

Control Number: H26326

Quantity: 2.0 Litres

The above product gave the following results on analysis:

Base line drift:

$\Delta mV/\Delta t$ on 440nm channel [Pass if <0.09 mV/min] = 0.056 mV/min

$\Delta mV/\Delta t$ on 570nm channel [Pass if <0.06 mV/min] = 0.003 mV/min

Chemical QC

Date: 10 August 2016

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.