

CERTIFICATE OF ANALYSIS

for

Ultra Ninhydrin Solution

Part Number: 80-2117-64

Control Number: H26387

Quantity: 250 mL

The above product gave the following results on analysis:

Base line drift:

$\Delta mV/\Delta t$ on 440nm channel [Pass if <0.09 mV/min] = 0.034 mV/min

$\Delta mV/\Delta t$ on 570nm channel [Pass if <0.06 mV/min] = 0.026 mV/min

Colorimetric test:

Absorption at 485nm [Neat, ref. To water. Pass if <0.10] = 0.045 A

Absorption at 600nm [Neat, ref. To water. Pass if <0.05] = 0.029 A

Chemical QC

Date: 19 September 2016

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.