



Biochrom Ltd
Certificate No. 890333

CERTIFICATE OF ANALYSIS

for

Ultrosolve Plus

Part Number: 80-2117-65

Control Number: H26388

Quantity: 1.75 Litres

The above product gave the following results on analysis:

pH = 5.629 @ 20.0 °C

pH Specification = 5.65 ± 0.10 @ 20.0 °C

Conductivity = 19.9 mScm-1 @ 20.0 °C

Conductivity Specification = 20.5 ± 3.0 mScm-1 @ 20.0 °C

Chemical QC

Date: 19 September 2016

Postal Address

Biochrom Ltd
1020 Cambourne Business Park
Cambourne
Cambridge CB23 6DW
England

Telephone

+44 (0)1223 423723

Telefax

+44 (0)1223 420164

Email:

support@biochrom.co.uk

Web Site: <http://www.biochrom.co.uk>

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.