

## CERTIFICATE OF ANALYSIS

for

### Sodium Hydro/Oxid Buffer 2

Part Number: 80-2038-08

Control Number: H26397

Quantity: 2.0 Litres

The above product gave the following results on analysis:

pH = 4.245 @ 20.0 °C

pH Specification = 4.25 ± 0.02 @ 20.0 °C

Conductivity = 14.1 mScm-1 @ 20.0 °C

Conductivity Specification = 14.25 ± 1.1 mScm-1 @ 20.0 °C

Chemical QC

Date: 28 September 2016

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		