

CERTIFICATE OF ANALYSIS

For

Sodium Hydrolysate Buffer 5

Part Number: 80-2038-09

Batch (Control) Number: H26419

Quantity: 2 Litres

Expiry Date: September 2019

The above product gave the following results on analysis:

pH = 6.447 @ 20.0°C

pH Specification = 6.43 - 6.47 @ 20.0°C

Conductivity = 78.2 mScm⁻¹ @ 20.0°C

Conductivity Specification = 75.2 - 82.4 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
Neil Booth

Date of Issue: 12 September 2016

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.