



Biochrom Ltd  
Certificate No. 890333

## CERTIFICATE OF ANALYSIS

for

### Sodium Hydrolysate Buffer 5

Part Number: 80-2038-09

Control Number: H26419

Quantity: 2.0 Litres

The above product gave the following results on analysis:

pH = 6.447 @ 20.0 °C

pH Specification = 6.45 ± 0.02 @ 20.0 °C

Conductivity = 78.2 mScm<sup>-1</sup> @ 20.0 °C

Conductivity Specification = 78.82 ± 3.57 mScm<sup>-1</sup> @ 20.0 °C

Chemical QC

Date: 12 October 2016

Postal Address

Biochrom Ltd  
1020 Cambourne Business Park  
Cambourne  
Cambridge CB23 6DW  
England

Telephone

+44 (0)1223 423723

Telefax

+44 (0)1223 420164

Email:

support@biochrom.co.uk

Web Site: <http://www.biochrom.co.uk>

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.